§ 880.6430 Liquid medication dispenser.

- (a) *Identification.* A Liquid medication dispenser is a device intended for medical purposes that is used to issue a measured amount of liquid medication.
- (b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807. The device also is exempt from the good manufacturing practice regulation in part 820, with the exception of §820.180, with respect to general requirements concerning records, and §820.198, with respect to complaint files.

§880.6450 Skin pressure protectors.

- (a) *Identification*. A skin pressure protector is a device intended for medical purposes that is used to reduce pressure on the skin over a bony prominence to reduce the likelihood of the patient's developing decubitus ulcers (bedsores).
- (b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807. The device also is exempt from the good manufacturing practice regulation in part 820, with the exception of §820.180, with respect to general requirements concerning records, and §820.198, with respect to complaint files.

§880.6500 Medical ultraviolet air purifier.

- (a) *Identification.* A medical ultraviolet air purifier is a device intended for medical purposes that is used to destroy bacteria in the air by exposure to ultraviolet radiation.
- (b) Classification. Class II (performance standards).

§ 880.6710 Medical ultraviolet water purifier.

- (a) *Identification.* A medical ultraviolet water purifier is a device intended for medical purposes that is used to destroy bacteria in water by exposure to ultraviolet radiation.
- (b) Classification. Class II (performance standards).

§880.6730 Body waste receptacle.

(a) *Identification*. A body waste receptacle is a device intended for medical

purposes tht is not attached to the body and that is used to collect the body wastes of a bed patient.

(b) Classification. Class I (general controls). The device is exempt from the premarket notification procedures in subpart E of part 807. The device also is exempt from the good manufacturing practice regulation in part 820, with the exception of §820.180, with respect to general requirements concerning records, and §820.198, with respect to complaint files.

§880.6740 Vacuum-powered body fluid suction apparatus.

- (a) Identification. A vacuum-powered body fluid suction apparatus is a device used to aspirate, remove, or sample body fluids. The device is powered by an external source of vacuum. This generic type of device includes vacuum regulators, vacuum collection bottles, suction catheters and tips, connecting flexible aspirating tubes, rigid suction tips, specimen traps, noninvasive tubing, and suction regulators (with gauge).
- (b) *Classification*. Class II (performance standards).

§880.6760 Protective restraint.

- (a) *Identification*. A protective restraint is a device, including but not limited to a wristlet, anklet, vest, mitt, straight jacket, body/limb holder, or other type of strap, that is intended for medical purposes and that limits the patient's movements to the extent necessary for treatment, examination, or protection of the patient or others.
- (b) Classification. Class I (general controls).

[61 FR 8439, Mar. 4, 1996]

§ 880.6775 Powered patient transfer device.

- (a) *Identification*. A powered patient transfer device is a device consisting of a wheeled stretcher and a powered mechanism that has a broad, flexible band stretched over long rollers that can advance itself under a patient and transfer the patient with minimal disturbance in a horizontal position to the stretcher.
- (b) Classification. Class II (performance standards).